MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN	PRINTER SN	DATE OF INSPECTION
09 7401	09.3580.879	7-31-13

LOCATION OF INSTRUMENT (STREET AND CITY)	TIME OF INSPECTION
600 E Walnut Columbia	1815

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

☑ DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

☑ TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

☑ PRINTER WORKING PROPERLY

☑ TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

☑ SIMULATOR SOLUTION   ☐ COMPRESSED ETHANOL-GAS MIXTURE

☑ STANDARD SUPPLIER Guth Labs   LOT # 13016   EXP. DATE 1-9-15

☑ SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0°C SIMULATOR SN SP 2761 SIMULATOR EXP DATE 7/2013

☑ CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
☑ 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
☐ 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
☐ 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ☞ .099	TEST 2 ☞ .100	TEST 3 ☞ .099

☑ RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	(OVER .19)
3	3	1	2	1	1

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE	PRINT NAME
[signature]	Nathan Turner

TYPE II PERMIT NUMBER/EXPIRATION DATE	TELEPHONE NUMBER
22411 12-20-14	573-874-7652

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

MO 580-1351 (6-10)   AN EQUAL OPPORTUNITY/AFFIRMATIVE ACTION EMPLOYER   LAB-114
services provided on a nondiscriminatory basis

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 13010 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on January 14, 2013, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1218% (w/vol) ethyl alcohol. The expiration date for this lot number is January 9, 2015 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:
Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.
All balances are calibrated annually by an outside agency using NIST traceable weights.
Calibration verification is done prior to each use utilizing NIST traceable weights.

AS IV Serial no: 097401
Version   no: 004C

TEST RECORD   01380

Temp   Date   Time   g/
                      210L
----   ----   ----   ----
Air Blank:
       07/31/13 18:15 .000
Calibration Check:
22   07/31/13 18:15 .099

Subject Name
    Test.
Subject I.D.

Operator Name, I.D.
    Turner 1585
Location
    CPP

AS IV Serial no: 097401
Version   no: 004C

TEST RECORD   01383

Temp   Date   Time   g/
                      210L
----   ----   ----   ----
Air Blank:
       07/31/13 18:22 .000
Calibration Check:
23   07/31/13 18:22 .099

Subject Name
    Test
Subject I.D.

Operator Name, I.D.
    Turner 1585
Location
    CPD

AS IV Serial no: 097401
Version   no: 004C

TEST RECORD   01381

Temp   Date   Time   g/
                      210L
----   ----   ----   ----
Air Blank:
       07/31/13 18:17 .000
Calibration Check:
22   07/31/13 18:17 .100

Subject Name
    Test
Subject I.D.

Operator Name, I.D.
    Turner 1585
Location
    CPP

AS IV Serial no: 097401
Version   no: 004C

TEST RECORD   01384

Temp   Date   Time   g/
                      210L
----   ----   ----   ----
Void: RFI
12   07/31/13 18:23

Subject Name
    Test
Subject I.D.

Operator Name, I.D.
    Turner 1585
Location
    CPP